



## Request for Quotation RfQ18/01731 Amendment no. 2

Ref. no. RfQ18-01731

Date: April 26, 2018

Subject: "Procurement of Medical Equipment for Ministry of Interior Hospital"

Dear Sir/Madam,

1. UNDP Moldova is hereby amending the solicitation document.
2. Due to the need to adjust some tender technical requirements, Annex 1, Lot 1 – M-01, Lot 2 M-02 are hereby amended to read as follows (changes are marked in red):

For both Lot 1 M-01 ENT Surgical and Lot 2 M-02 ENT Diagnostic:

- |     |                   |                                   |
|-----|-------------------|-----------------------------------|
| 2-6 | Video camera head |                                   |
| (1) | Type              | : HD                              |
| (2) | Applicable        | : Applicable for 1-1(1), (2), (3) |
| (3) | Controls          | : Programable control buttons     |
| (4) | Zoom              | : Optical                         |
| (5) | Camera chips      | : 1 or 3                          |

For Lot 1 M-01 ENT Surgical:

- |      |                           |                                |
|------|---------------------------|--------------------------------|
| 2-11 | Suction - irrigation unit |                                |
| (1)  | Controls                  | : Equipped, pedal or hand type |
| (2)  | Irrigation Pump           | : equipped                     |
| (3)  | Irrigation pump capacity  | : 2 to 3 l/min                 |
| (4)  | Suction pump capacity     | : 2 to 3 l/min                 |
| (5)  | Pressure                  | : 350-400 mm Hg                |
| 2-13 | Motor system for shaver   |                                |
| (1)  | Motor speed               | : 30.000 - 40.000 rpm          |

- (2) Handpiece speed : 15000 rpm or more
- (3) Display : 5 inches or more
- (4) Control : Foot switch

- 3. All other terms and conditions of the solicitation documents, except as amended herein, shall remain unchanged and shall continue in full force and effect.