

Request for Quotation RfQ18/01830 Amendment No. 1

Ref. no. **RfQ18/01830** Date: **24 October 2018**

Subject: SCBM Health/Vacuum-assisted fetal delivery device

Dear Sir/Madam,

1. Due to the need to correct one of the Evaluation Criteria indicated in the Request for Quotation document (page 4), as well as some of technical specifications of the devices to be purchased indicated in Annex 1, the present RfQ as well as its relevant Annexes are hereby amended to read as follows (changes marked in RED).

2. All other terms and conditions of the solicitation document, except as amended herein, shall remain unchanged and shall continue in full force and effect.

Evaluation Criteria	☐ Technical responsiveness/Full compliance to requirements (according to Annex 1) and lowest price¹		
	☑ Minimum 3 (three) years of experience in delivering medical equipment/devices;		
	☑ Availability of certificates of quality and origin for the offered goods (EC andFDA Declaration of Conformity);		
	☑ Availability of Quality Certificates (ISO, etc.) of the items provided by the manufacturer;		
	☐ Maximum delivery period not to exceed 30 calendar days u signature of PO Contract;		
	☑ Warranty for a minimum period of 3 (three) years (according to the requirements of the technical specifications);		
	☑ Full acceptance of the Contract General Terms and Conditions.		

¹ UNDP reserves the right not to award the contract to the lowest priced offer, if the second lowest price among the responsive offer is found to be significantly more superior, and the price is higher than the lowest priced compliant offer by not more than 10%, and the budget can sufficiently cover the price difference. The term "more superior" as used in this provision shall refer to offers that have exceeded the pre-determined requirements established in the specifications.

Technical Specifications

ltem	Goods	Specification	UNIT OF MEASURE	QUANTITY
1.	Vacuum- assisted fetal delivery device	Purpose: The device is used to assist delivery of a baby. Specifications: Maximum flow: 30 L/min or more Maximum vacuum: 60 kPa or more Noise level: 60 dB or less Voltage: ±5%, calculated at 230V /50 Hz Collection jars (autoclavable): ≥1 L Selector knob for double collection jar exchange: yes	piece	8
		Trolley: yes Hydrophobic PTFE bacteria filter: 99,95% particle retention and 99,99% bacterial challenge efficiency On/Off Switch: yes High Vacuum Gauge including a display: 0- 100kPa / 0-760 mmHg On/Off foot-operated pedal: yes Suction pressure-regulating knob: yes PVC suction tube length: 3m or more Power: 110 W or more		
		Accessories: Silicon cups set for fetus vacuum extraction: minimum 2 cups of 50 mm and 2 cups of 60 mm, multiuse Any other accessory which is required for installation and operation should be included and described on your offered specification		
		General Requirements: Compliant with MDD 93/42/EEC or US FDA Regulations Manufacturer ISO 13485 Warranty at least 36 months		
		Training: Training for all laboratory personnel who will be using this item, on each individual site Training in Russian language		
		Documentation: User's manuals, preferable in Russian language (optional English) Maintenance manual, preferable in Russian language (optional English)		

*) NOTES:

- Any other additional components and accessories necessary for proper installation and functioning of the manikin should be included and described in the offer;
- All works for assembly of the manikin will be supported by the supplier;

If the goods are supplied together with a software, it shall be provided in Russian language.